

Refrigerator Temperature Log Month/Year _____ Storage Unit _____ Provider _____ VFC# _____

Routine Temperature Monitoring– Record:

- Time of day and Initials of staff recording temperatures
- Whether the LED is green (Y or N)
- **AM:** Record current temperature with an X, and minimum and maximum temperatures each with an M.
- **PM:** Record current temperature with an X.

Out-of-Range Response: If temperatures are in the shaded zone or the Data Logger LED light is red:

- DO NOT USE or DISCARD the vaccine.
- Obtain a Vaccine Incident Report and follow Steps 1–6.
(www.immunization.mt.gov under “VFC” link)
- Call the Montana Immunization Program at 444-5580.

Retain temperature logs for three years.

Day of Month		1		2		3		4		5		6		7		8		9		10		11		12		13		14		15	
Time of Day		am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm
Staff Initials																															
LED Green?																															
≥50°																															
49°				DO NOT USE vaccine that has been stored at these temperatures (shaded grey). See instructions at top of page.																											
48°																															
(C)	47°																														
8° 7° 6° 5° 4° 3° 2°	Refrigerator (F)	46°																													
		45°																													
		44°																													
		43°																													
		42°																													
		41°																													
		40°																													
		39°																													
		38°																													
		37°																													
36°																															
35°																															
34°																															
33°				DO NOT USE vaccine that has been stored at these temperatures (shaded grey). See instructions at top of page.																											
32°																															
≤31°																															

Refrigerator Temperature Log Month/Year _____ Storage Unit _____ Provider _____ VFC# _____

Routine Temperature Monitoring– Record:

- Time of day and Initials of staff recording temperatures
- Whether the LED is green (Y or N)
- **AM:** Record current temperature with an X, and minimum and maximum temperatures each with an M.
- **PM:** Record current temperature with an X.

Out-of-Range Response: If temperatures are in the shaded zone or the Data Logger LED light is red:


- DO NOT USE or DISCARD the vaccine.
- Obtain a Vaccine Incident Report and follow Steps 1–6.
(www.immunization.mt.gov under “VFC” link)
- Call the Montana Immunization Program at 444-5580.

Retain temperature logs for three years.

Day of Month	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Time of Day	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm
Staff Initials																
LED Green?																
≥50°																
49°																
48°																
(C) 47°																
8° 46°																
7° 45°																
6° 44°																
5° 43°																
4° 42°																
3° 41°																
2° 40°																
39°																
38°																
37°																
36°																
35°																
34°																
33°																
32°																
≤31°																

DO NOT USE vaccine that has been stored at these temperatures (shaded grey). See instructions at top of page.

DO NOT USE vaccine that has been stored at these temperatures (shaded grey). See instructions at top of page.



Vaccine Storage Unit Trouble-Shooting Log

Record issues with storage unit and corrective actions taken in the table below.
(e.g., near out-of-range temperatures, malfunctions, repairs)

Date	Time	Description of Issue	Corrective Action Taken	Staff Initials